

# Manufacturer's Declaration

**Company**            **Endress+Hauser Conducta GmbH+Co. KG**  
**Dieselstrasse 24, 70839 Gerlingen, Germany**

declares as manufacturer of analytical products that the process media wetted FKM-O-ring of our product

**Product**            **Memosens CPS76D**

**Regulations**        comply with the FDA regulation 21 CFR 1777.2600 and has passed the bioreactivity tests in vitro and in vivo according to USP <87> and USP <88> Class VI (121 °C).

Gerlingen, 31.07.2019  
Endress+Hauser Conducta GmbH+Co. KG

  
i. V. Uwe Rößiger  
Manager Certifications and Approvals

  
i. V. Dr. Angela Eubisch  
Expert Applied Biological Sciences