



Certificate of Compliance

Certificate: 1642298

Master Contract: 200600

Project: 1642298

Date Issued: 2005/05/10

Issued to: Endress + Hauser Wetzer GmbH Co. KG

Obere Wank 1
Nesselwang, 87484
Germany
Attention: Michael Pfanzelt

The products listed below are eligible to bear the CSA Mark shown



Issued by:

Edward Foo, C.E.T.

Authorized by: Nick Alfano, Operations
Manager

PRODUCTS

CLASS 2252 03 - PROCESS CONTROL EQUIPMENT

- Process Display Transmitter, Model RIA 250 Series, panel mounted rated 90-253 Vac, 50/60Hz, 11.5VA, 20-28Vac, 50/60Hz, 5.5VA or 18-36Vdc, 5.5 VA. Ambient temperature - 10°C to + 50°C. Pollution Degree 2, Installation Category II environment.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 1010.1-92 - Safety requirements for electrical equipment for measurement,

Amends 1 and 2 control and laboratory use.



Supplement to Certificate of Compliance

Certificate: 1642298

Master Contract: 200600

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
1642298	2005/05/10	CSA Certification of Process Display with UL Report 03CA14945 for General Purpose