

Certificate of Compliance

Certificate: 1135330

Master Contract: 200600

Project: 1378193

Date Issued: November 7, 2002


Issued to: **ENDRESS + HAUSER WETZER GMBH CO. KG**
Obere Wank 1
D-87484 Nesselwang
GERMANY

Attention: **Mr. Thomas Kellermann**

The products listed below are eligible to bear the CSA Mark shown



Issued by:


E. Foo, C.E.T.

Authorized by:


Nick Alfano
Operations Manager

PRODUCTS

CLASS 2252 01 - PROCESS CONTROL EQUIPMENT

Preline Process Transmitter, Model RMA 422, rated 90-250V ac, 11VA, 50/60Hz or 18-36V dc, 11VA/20-28V ac, 11VA, 50/60Hz with optional output relay contacts rated 250V, 5A, 50/60Hz.

Note: The above panel mounted equipment is Certified for use with other equipment in a suitable enclosure where the suitability of the combination is to be determined by CSA or Local Inspection Authority.

Certificate: 1135330
Project: 1378193

Master Contract: 200600
Date: November 7, 2002

APPLICABLE REQUIREMENTS

CSA Std C22.2 No.142-M1987 - Process Control Equipment

MARKINGS

The following markings appeared on a Certified nameplate in a permanent and visible manner:

- Submitter's name or tradename and/or CSA Master Contract No 200600;
- Model designations;
- Electrical ratings;
- CSA Monogram;
- Serial number or date code.

Supplement to Certificate of Compliance

Certificate: 1135330

Master Contract: 200600

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
1378193	Nov. 7, 2002	Update to cover deletion of model suffix "D".
1135330	Jun. 11, 2001	Original Certification.