



CSA INTERNATIONAL

Certificate of Compliance

Certificate: 1769251

Master Contract: 200600

Project: 1769251

Date Issued: 2006/05/30

Issued to: Endress + Hauser Wetzer Gmbh Co. KG

Obere Wank 1
Nesselwang, 87484
Germany
Attention: Mr. Michael Pfanzelt


The products listed below are eligible to bear the CSA Mark shown



Issued by:


Edward Foo, C.E.T.

Authorized by: Patricia Pasemko, Operations
Manager



PRODUCTS

CLASS 2252 05 - PROCESS CONTROL EQUIPMENT

· Multi channel Recorder Ecograph T, panel-mounted, model RSG30-***, rated 115 - 230Vac, 50/60Hz, 30VA; 24Vac, 50/60Hz, 24VA or 24Vdc, 24VA. Output rated 28Vdc, 250mA and relays contacts ratings 50Vdc, 300mA or 230Vac, 3A. Ambient temperature 0°C to 50°C. Pollution Degree 2, Installation Category II environment.

Notes:

- For use in other equipment where the suitability of the combination shall be determined by CSA;



CSA INTERNATIONAL

Certificate: 1769251

Master Contract: 200600

Project: 1769251

Date Issued: 2006/05/30

- The 24Vdc and 20-28Vac supplies must be powered by SELV supply source.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety requirements for electrical

Second Edition equipment for measurement, control and laboratory use, Part 1: General Requirements



CSA INTERNATIONAL

Supplement to Certificate of Compliance

Certificate: 1769251

Master Contract: 200600

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
1769251	2006/05/30	CSA Certification of Ecograph T for Ordinary Locations Classification