

Declaration of Compliance

Endress+Hauser GmbH+Co. KG, Hauptstraße 1, 79689 Maulburg

being the manufacturer, declares that the product stated below

Ceraphant PTC31-A4C11H1AE1A


has passed the listed basic quality and functionality tests:

	<i>Activity</i>	<i>Test specification</i>	<i>Test criteria</i>	<i>Frequency of testing</i>
1	Contract review	Internal test specification TS00001F/00/EN	As required in ISO 9001	100%
2	Incoming goods inspection	Internal test specification TS00003F/00/EN	Verification of conformance to the specified requirement	Dynamic sample rate according to CAQ (Computer aided Quality System)
3	Printed circuit boards test	Internal test specification TS00009F/00/EN	In-circuit test and functional test	100%
4	Calibration of instrument	Internal test specification for pressure and differential pressure measurement TS00023P/00/EN	Measurement, adjustment and verification of lower range value, upper range value and output signal	100%
5	Final check of instrument	Internal test specification for pressure and differential pressure measurement TS00024P/00/EN	Verification of general functionality and customer specific settings	100%
6	Outgoing visual check	Internal test specification TS00006F/00/EN	Visual inspection of completeness and correctness of the instrument and the markings	Partial inspection
7	Logistical verification	Internal test specification TS00005F/00/EN	Automatic verification in the delivery phase via barcode that the instrument correspond to the order	100%


Order No.	Order Code	Serial No.
3018667775	PTC31- A4C11H1AE1A	LB017A01999
		LB017B01999
		LB017C01999
		LB017D01999

This declaration of compliance is exclusively valid for the customer listed in the cover letter of the respective Endress+Hauser sales center and for the listed products and accessories in delivery status.

Maulburg, 20-December-2016
Endress+Hauser GmbH+Co. KG



i. V. Dr. Arno Götz
Dept. Manager Product Safety
Research & Development Platform



i.V. Gerd Gritsch
Division Manager Quality
Management